



# DISCONTINUED: VP2 Universal Anemometer Interface

## Product Images



## Short Description

---

With this interface, some anemometers not manufactured by Davis Instruments can be used.

## Description

---

With this interface, some anemometers not manufactured by Davis Instruments can be used with the Vantage Pro2. Non-Davis anemometers must have AC voltage output signals. The frequency of the AC signal must be linearly proportional to the wind speed.

## Additional Information

---

Brand	Davis Instruments
Promotion	<b>This product has been discontinued. See the <a href="#">DAVIS WIRELESS SOLAR-POWERED SENSOR TRANSMITTER</a></b>
Explanation	This interface has been tested with R. M. Young anemometers, model number 05103; and with NRG anemometer, model 1900 NRG #40C. However, the interface is not limited to those products.
Ideal For	Professional